

# **Vaginal Self-Swab Device to Minimize Contact Contamination**

**Client:** Dr. Jean Riquelme

**Advisor:** Prof. Randolph Ashton

**Date:** 09/30/2024

## **Team:**

Jackie Behring	jbehring@wisc.edu (Co-Team Leader & BSAC)
Tatiana Predko	tpredko@wisc.edu (Co-Team Leader & Communicator)
Mariah Smeeding	jimenezsmeed@wisc.edu (BWIG)
Ava Fevold	fevold@wisc.edu (BPAG)

## **PROBLEM STATEMENT**

Bacterial Vaginosis, yeast infections, and sexually transmitted infections (STIs) can be detrimental to the well-being of an individual and cause a variety of health concerns if left untreated. The vaginal self-swab device is to be utilized by patients to easily collect cervicovaginal mucus samples from the vaginal canal to diagnose vaginal infections and STIs. This device design aims to provide a convenient, accessible method of breaking the swab into the transfer tube while minimizing cross-contamination of the self-collected sample. Cross-contamination, with the surface and environment, typically occurs while transferring the sample to the culture media, which can alter the results. In order to overcome this, the device should allow the testing swab to break into the culture media solution directly, and to prevent media leakage.

## **BRIEF STATUS UPDATE**

- The team worked on completing the preliminary presentation. Meetings were scheduled between team members on 9/29 and 10/2 to assign roles and speaking parts to the presentation slides.

## **SUMMARY OF WEEKLY TEAM MEMBER DESIGN ACCOMPLISHMENTS**

- Team:
  - Completed the Preliminary Design presentation slides.
  - Met as a team to practice presenting.
  - Met as a team over zoom to assign slides for the preliminary presentation.
- Jackie Behring:
  - Met as a team to assign speaking parts and slides for presentation
  - Completed assigned preliminary presentation slides
  - Met as a team to practice for presentation
- Tatiana Predko:

- Met with the team via Zoom to assign preliminary presentation roles.
- Met with the team to practice presenting.
- Completed additional research on competing designs and added into LabArchives.
- Mariah Smeeding:
  - Met with team to rehearse preliminary presentation
  - Completed the given portion of the preliminary presentation and rehearsed
  - Submitted progress report 4 and preliminary presentation to our website.
- Ava Fevold:
  - Worked on the Preliminary Design Presentation
  - Continued to research and update LabArchives
  - Rehearsed Preliminary Design Presentation with group

### **PROJECT DIFFICULTIES**

- None

### **UPCOMING TEAM AND INDIVIDUAL GOALS**

- Team:
  - Begin working on preliminary report
  - Meet with the client to update her on the design process and address any concerns
- Jackie Behring:
  - Begin modeling the chosen design in SolidWorks
  - Complete preliminary report assigned sections
  - Continue researching and keep lab archives updated
- Tatiana Predko:
  - Email the client and advisor with updates on the project, and confirm the next meeting time.
- Mariah Smeeding:
  - Continue researching possible contaminants to the self sampling process.
  - Present preliminary presentation
  - Meet with client
  - Add further research to lab archives
- Ava Fevold:
  - Complete assigned sections of the preliminary report
  - Update Lab Archives with any recent research
  - Meet with client and team




### **PROJECT TIMELINE**

<b>Task</b>	<b>Sep.</b>	<b>Oct.</b>	<b>Nov.</b>	<b>Dec.</b>



TOTAL:	\$0.00							

## DESIGN MATRIX

Design Criteria (weight)	Design 1: Altered Bend Design		Design 2: Tilt - and - Break		Design 3: The Tunnel	
						
Limiting Contamination (30)	4	24	4	24	4	24
Leakage Prevention (25)	4	20	5	25	4	20
Ease of Use (15)	4	12	5	15	3	9
Ease of Fabrication (10)	3	6	4	8	2	4
Patient Comfort (10)	5	10	5	10	5	10
Safety (5)	4	4	5	5	2	2
Cost (5)	5	5	5	5	5	5
Total Score (100)	81		92		74	

### Determination of Criteria and Weights:

- Limiting Contamination (30): Leaking contamination refers to the ability to inhibit the spread of microorganisms from biological fluids or the testing environment. This criterion was discussed in the *Accuracy and Reliability* section of the PDS, and was assigned a weight of 30/100 due to its significance of being an accurate measurement. The device must demonstrate a consistent snapping mechanism while transporting the swab into the culture media. This requires the minimization of contamination caused by

the patient's own bacterium, such as their hands, or from the environmental bacterium, such as a countertop.

- Leakage Prevention (25): Leakage prevention refers to the degree to which spilling of contained culture media is prevented. This criterion was assigned a weight of 25/100 because it is of paramount importance that the culture media does not leak. Leakage of the contained culture media can lead to inaccurate testing results, due to partial loss of the sample or external contamination. Additionally, skin contact with the culture media should be avoided, as it could lead to injury of the patient.
- Ease of Use (15): This section focused on the simplicity of the process using the device. Relevant information was covered in the *Life in Service* and *Shelf Life* sections of the PDS. This criterion was given a weight of 15/100 because it is important that a patient would be able to understand and correctly use the device. A simple testing process is essential since the self-swab would be done without a doctor or physician's guidance.
- Ease of Fabrication (10): This criterion assesses the ease of fabricating the insertion and stand mechanism, including the production and assembly of its components. Ease of fabrication was given a 10/100 because it's crucial that the design is feasible within the given time frame, both for prototyping and potential commercial manufacturing. A higher score in this category reflects a design that aligns with our fabrication skills and can be more readily scaled for commercial production.
- Patient Comfort (10): Patient comfort refers to the comfort level of the patient while using the product, notably to which degree discomfort and pain are avoided. This criterion was assigned a weight of 10/100, as it is not an essential component of device functionality, but it is vital in the marketability and convenience of the product. Due to its convenience, patients are more likely to self-swab than to obtain a physician-collected sample [1].
- Safety (5): Safety refers to risks (specifically to the user) associated with the product. This category ensures that the shaft portion of the swab can only be inserted 5 cm to prevent mechanical-induced tissue injury to the user [2]. Every material to be used is known to be biocompatible with the human body to reduce the risk of hypersensitivity [3].
- Cost (5): The cost portion of the matrix is used to determine whether the cost of the materials are justifiable and within the client's budget. Points off in this category indicate costs exceeding the budget.

### **Justification of Assigned Scores:**

- Limiting Contamination (30): The "Altered Bend" design scored a 4% because although the base was extended to provide more stability, the device still has the potential of being knocked over, which could lead to contamination if the swab comes into contact with an external surface. The "Tunnel" design scored a 4% in limiting contamination because the

mechanism is made out of malleable plastic, meaning it could tip easier without the added weight. Similarly to the other two design concepts, the “Tilt-and-Break” design received a score of 4. This is because although the design consists of a solid “handle” (as compared to the hollow “handles” of the other two designs) making it less prone to tipping over, there is still the possibility of the device being knocked over and becoming contaminated after unintentionally coming into contact with a surface.

- Leakage Prevention (25): The “Altered Bend” design scored a 4 in the leaking prevention category because of its inability to screw onto the device containing the transportation medium. The patient could accidentally knock over the tube containing the medium, leading to the potential spilling of viable bacterium through the crevices of the device. Similarly to the “Altered Bend” design, there is a possibility for tipping with the “Tunnel” design, so it is ranked a 4. Since there is no solid divide between the swab-holder and the tube, the culture media could spread throughout the hollow holder and escape the container. The “Tilt-and-Break” design scored full points in this category, because the “handle” component of the device is solid, and has no hollow voids for fluid to seep into if the device is not upright. Additionally, this design utilizes threads to allow for a secure, “screw-on” tightening of the two components.
- Ease of Use (15): The “Altered Bend” design was given a score of 4 for ease of use because the button mechanism could malfunction, leading to inconclusive results. The “Tunnel” design was ranked the lowest, with a 2, because of the need to apply substantial, and variable (due to the user), pressure in order to break the swab. Finally, the “Tilt-and-Break” design scored full points in this category. This is because the device solely consists of a simple “screwing-on” mechanism, making it simple to use.
- Ease of Fabrication (10): When it comes to the ease of fabrication, the “Altered Bend” design was assigned a score of 4 because, in previous semesters, the button had issues with rotating, providing an inaccurate breaking of the swab (either too long or too short). Fabricating a “button” mechanism and assembling it in a non-rotational manner would be difficult. Ranked 2, the “Tunnel” design would be difficult to fabricate due to the incorporation of sharp inserts within the holder used to stabilize the swab. Lastly, the “Tilt-and-Break” design was given a score of 4; this device would be relatively easy to fabricate due to its simplistic design, but it may be difficult to create a “handle” into which the swab can be sufficiently pressure-fit. Additionally, the “Tilt-and-Break” design utilizes threads, which is an added component of difficulty with regard to the ease of fabrication.
- Patient Comfort (15): Each of the designs scored full points for patient comfort because all of the designs have a similar mechanism in terms of specimen collection by the patient. During the process, the patient must insert the swab 5 cm into the vaginal canal, which provides little-to-no discomfort to the patient if done correctly [2].
- Safety (5): The “Altered Bend” design received a score of 4 for safety because the button could potentially malfunction and pinch the user if not pressed correctly. The “Tunnel”

design ranked the lowest, with a  $\frac{2}{5}$ , due to the sharp insert at the bottom of the holder. This can be dangerous to the user during the swabbing process, as well as while breaking the swab. The simplistic design of the “Tilt-and-Break” mechanism leaves little room for the user to injure themselves, which is why this design received a full score in this category.

- Cost (5): Each of the designs received full points within the cost category because all of the proposed design ideas require a minimal amount of material for fabrication, leading to affordability. All of the costs required to manufacture either device are anticipated to be within the \$250 budget of the client.