

# **Scale to Measure Nighttime Weight Change**

Kyle Herzog – Team Leader  
Joe Ferris – BWIG  
Ben Engel – Communications  
Eric Printz – BSAC

Advisor: Wally F. Block, Professor of Biomedical Engineering  
Client: Timothy Juergens, M.D., Geriatric psychiatrist/sleep specialist

October 15, 2006

## Table of Contents

<b>TABLE OF CONTENTS .....</b>	<b>2</b>
<b>ABSTRACT.....</b>	<b>3</b>
<b>MOTIVATION .....</b>	<b>4</b>
<b>PROBLEM STATEMENT.....</b>	<b>4</b>
<b>CLIENT REQUIREMENTS.....</b>	<b>4</b>
<b>BACKGROUND INFORMATION .....</b>	<b>5</b>
<i>Correlation between Obesity and Weight Gain Before Bed.....</i>	<i>5</i>
<i>Correlation between Obesity and Insomnia.....</i>	<i>7</i>
<i>Analog to Digital Converter.....</i>	<i>8</i>
<i>Amplifiers.....</i>	<i>9</i>
<b>ALTERNATE DESIGNS.....</b>	<b>9</b>
<i>Load Cell .....</i>	<i>9</i>
<i>Pad Design.....</i>	<i>11</i>
<i>Tension Meter.....</i>	<i>13</i>
<b>DESIGN MATRIX .....</b>	<b>16</b>
<b>FUTURE WORK.....</b>	<b>17</b>
<b>PROBLEMS AND POTENTIAL SOLUTIONS.....</b>	<b>18</b>
<b>REFERENCES.....</b>	<b>19</b>
<b>APPENDIX.....</b>	<b>20</b>

## **Abstract**

It has been proposed to design a weighing device that can measure weight change of a patient in a hospital bed over the course of night. The weight of the patient should be recorded every ten to fifteen minutes throughout the night. The device would likely have to support a hospital bed sitting on four legs and ideally read at an accuracy around 0.02-0.05 lbs. The more accurate the device the more useful it will be. The device should be capable of being moved between beds with minimal difficulty and without any complex recalibration required.

## **Problem Statement**

For most people, weight fluctuates throughout the day and night. As a result one can lose as much as 700 grams over the course of the night, even if they do not eat or go to the bathroom. Much of this is related to metabolism, as well as water loss. We would like to be able to measure this weight change accurately on people who are in hospital beds. This will require that our device record the patient's weight every fifteen minutes. The device would likely have to support a hospital bed, which sits on four legs, and read at an accuracy between 0.02- 0.05 lbs. Ideally the device could be moved to a different bed without too much difficulty (one to two hours time) and without complex recalibration.

## **Motivation**

Tracking weight change over the course of a night may be important from a research perspective. There are a number of potential variables that may affect this weight change, such as medications and sleep stage changes. Another potential use would be to track and eventually predict those at higher risk of developing obesity, correlate with body fat measurements, activity, and assess for metabolic changes when we do things to improve or correct sleep problems. There is popular concern in recent sleep literature regarding various sleep stages, insomnia, and their impact on obesity.

## **Client Requirements**

One of the device's important client requirements is that it output weight readings to a computer at chosen time intervals. The user should be able to choose the frequency of readings and have the data values imported to a program that is as universal as possible, such as Microsoft

Excel. This will allow for easy analysis and manipulation of the data. The more accurate the scale, the more useful the readings will be. An accuracy of 0.02 to 0.05 lbs is desired, but a slightly less precise scale would still be very useful. Achieving accuracy this high with materials that are of reasonable cost will be very difficult and a lower accuracy may be used in the prototype designed this semester. Another important client requirement is that the scale not be fixed to one bed. It should be capable of being moved from one bed to another without the need for many hours of labor or complex recalibration.

### **Background Information**

This semester is the first in which this specific project has been worked on by a group of biomedical engineering students. Dr. Juergens provided us with several journal articles in our first meeting that help to explain the motivation for such a device and the possible benefits it could provide. These papers, in conjunction with some other outside research, provided us with a wealth of information on which to base our designs.

#### *Correlation between obesity and weight gain before bed*

A recent study performed at Nakamura-Gauken University in Japan proved that there is a direct correlation between the daily fluctuation in body weight immediately before going to bed and weight regain and accompanying visceral fat accumulation in obese patients. The study consisted of a 16 month weight-reduction program with an initial four month education program. The 16 month weight-reduction program that the study focused on was completed by ninety-eight women in a charting group (recorded weights at four different times of day) and seventeen women in a noncharting group (no weight measurements). Consequently, ninety-eight obese

women who ranged in age from 23 to 66 years who had no history of major diseases or medications known to influence the parameters examined in the study were assessed. Weight recordings were made at four separate times during the day: immediately after waking up, after eating breakfast, after dinner, and before going to bed. Abdominal fat areas and blood samples were assessed at 0, 4, and 16 months. To ensure the comparability of the charts between days, adherence to the exact times and weighing instructions was critical. The study found that plotting the daily body weight measured at four critical times each day is effective in weight reduction, maintaining weight loss, and preventing subjects from dropping out of a weight loss program. Based on the rate of attrition, the performance of subjects in the charting group was clearly superior to that in the noncharting group, which confirms previous findings.

The results also revealed that the standard deviations of the difference in body weight between “after waking up” and “before going to bed” predicted the rebound increase in body weight. Patients who had an inconsistent weight before going to bed (higher than the weight after dinner) quickly regained their lost weight by the end of the 16 months study. Obese patients have the distinct characteristic of excessive snacking during the period before going to bed, which is called *night-eating syndrome*. Excessive energy intake during the night results in a rebound increase in fat accumulation. When compared, the study’s charting results imply that an irregular weight pattern during the period before going to bed is a sign of excessive energy intake during the night (4).

The device we are designing would open the door to a variety of studies similar to the one above. An obese patient’s weight change over the course of a night could be correlated to his or her overall weight regain. For example, data obtained using such a scale may show that there is excessive weight regain during the REM stages of sleep. It may then be possible to

prevent or minimize this weight gain by altering the sleep cycle through medication or some other type of intervention. The study performed at the Nakamura-Gakuen University implies that weight gain which occurs while the test subject is asleep definitely has a direct affect on the amount of weight a patient regains over a longer period of time. Studies using a scale similar to the one we are designing could help to further explain this correlation.

### *Correlation between Obesity and Insomnia*

In a recent study performed by Yale School of Medicine researchers showed that there is a possible link between insomnia (lack of sleep) and obesity. The link between the two has been traced to hypocretin/orexin cells in the hypothalamus region of the brain that are easily excited and sensitive to stress. The researchers found that if these cells are over-activated by environmental or mental stress in daily situations, they may support sustained arousal, triggering sleeplessness which leads to overeating. The more stress that a person undergoes the lower the excitation threshold of these hypocretin cells becomes. The study was performed by using electrophysiology and electron microscopy to observe hypocretin/orexin neurons in mice. Researchers found a previously undescribed organization of inputs on hypocretin neurons in which excitatory nerve junctions outnumber inhibitory contacts by almost ten fold. This unique wiring correlates well with their involvement in the control of arousal and alertness, but may also be an underlying cause of insomnia and associated metabolic disturbances, including obesity.

This study proves that there is in fact a direct correlation between obesity and certain sleep disorders. A functioning scale that tracks weight change over the course of the night could help Dr. Juergens and his colleagues to explain what it is about these sleep disorders that leads to a higher prevalence of obesity (5).

## *Analog-Digital Converter*

An analog digital converter will most likely be needed to convert the voltage being produced by the measuring device to a discrete number. Traditionally an analog digital converter, or ADC, receives a continuous signal output and produces discrete values that can be output digitally. There are many factors to consider when choosing the correct ADC for your task.

The resolution of the converter indicates the number of discrete values it can produce over the range of voltage values. The voltage resolution of an ADC is equal to its overall voltage measurement range divided by the number of discrete values. The resolution required would vary depending on the voltage range of the measuring device used. The resolution is limited by the signal-to-noise ratio of the signal being input. If there is too much noise in the incoming analog signal, it will be impossible to resolve the signal beyond a certain number of bits of resolution. The number of bits that are capable of being resolved is known as the "effective number of bits" (ENOB). While the ADC will produce a result even if the signal-to-noise ratio is very low, the result is not accurate, since its lower bits are simply measuring noise. The ratio should be around 6 dB per bit of resolution required.

Non-linearity may also be a problem. All ADCs suffer from some non-linearity as a result of their physical imperfections. This can cause incorrect readings that can skew data. This can be minimized by properly calibrating the ADC before it is put into use, or by testing to discover an approximate value of this non-linearity and then compensate for it accordingly (3).



## *Amplifiers*

An amplifier will be required to amplify the signal being produced by the measuring device to a level that can be recognized by the ADC. An amplifier simply uses a small amount of energy to control a larger amount of energy. Most amplifiers can be characterized by several parameters. The extent to which a signal is amplified is called gain and is usually measured in decibels (dB). It is defined very simply as the output level divided by the input level. Noise is frequently introduced in the amplification process and must be taken into account. It is also measured in decibels and is an undesired but inevitable effect of the electrical set up and components.

The efficiency of an amplifier is a measurement of how much of the power being put into the amplifier is actually being used for amplification. An inefficient amplifier would require a very high input of power, but would produce a very low level of signal amplification; therefore, a high level of efficiency is desired. As with ADCs, a certain amount of non-linearity also arises at specific signal input levels. An ideal amplifier would be a completely linear device, but amplifiers are actually only linear over a certain range that is dependent upon the amplifier itself (2).

## **Design Alternatives**

### *Load Cell*


A load cell is a device that is used to measure the weight of an object based on the fact that the resistance of a material will change as the load applied to the material changes. When a load cell is subjected to a force, the change in resistance is related to the magnitude of the force applied to it. Ultimately the change in resistance is converted into a weight. In this design we

would use four strain gauge load cells, each under one leg of the bed. Each load cell will be connected to an amplifier circuit which amplifies the signal before sending it to a computer for interpretation and logging (1).

The key component between the amplifying circuit and the computer will be an analog/digital converter. This device takes an analog electrical signal from the amplifier (which is connected to a load cell), and converts it into a serial signal which the computer can interpret. Therefore, our device will connect to the computer using the computer's serial jack. Since most computers have only one serial port, an additional PCI card may be needed to expand the computers serial port capacity to four (See Figure 1).

The computer will interpret the signal and put the data into an excel spreadsheet, as our client has requested. Some analog/digital converters come with data recording software. If this software is able to record our weight data in fifteen minute intervals in a way such that the data is useful to our client, then we will use this software. If the included software does not have sufficient features then our second option is to write our own software using ActiveX controls.

Overall the load cell design option has many advantages. Load cells are currently the



**Figure 1 – Load Cell Design**

Note: While only two bed posts are shown here, the final design will include a load cell (with the same components between it and the computer) for each of the legs that aren't shown.

“state of the art” way to weigh objects of a wide variety of sizes. Through our research we have found that a fairly standard accuracy for a load cell is 0.02% of its full load. We feel that this is a sufficient accuracy because our load will be spread out among four load cells. Spreading out the load will help to increase accuracy. Furthermore, this design will be versatile. As new load cell technology is developed it will be fairly easy to purchase a new load cell and hook it up to the amplifying circuit. Also, from a data logging standpoint, we will be able to record and interpret the data in just about any way because of the ability to write our own programs that interface with the analog/digital converter.

The main disadvantage of this design is cost. The average price of a load cell is around \$300. We will need four of these. The amplifying circuit and analog/digital converter will cost less than \$40 so those aren't as much of a concern. To remedy this problem we would purchase only one low capacity load cell for under \$40. If we are able to record data from this load cell on a computer, then we feel that we would be confident enough to purchase four higher capacity load cells. Another disadvantage is portability. Since there are four load cells with multiple components between each of them and the computer, the device may be difficult to move for someone who isn't familiar with the device.


### ***Pad Design***

Weighing pads are a type of scale that are very thin and have a large surface area. They are generally made out of stainless steel and come with a rubber mat for comfort. Weighing pads utilize a digital LCD readout and are most often powered by either AC adapters or a variety of different batteries. They are commonly used by veterinarians to weigh animals. In our weighing pad design we would place one of these pads under the mattress of the bed. Data is transferred to

LCD readouts via a parallel connection. We would need to reroute the wiring on this parallel connection and input the data into a computer system through a parallel port. The computer system would be designed to take readings at a designated interval, recording throughout the course of the night, preferably using an Excel spreadsheet (See Figure 2).

The weighing pad design offers a number of advantages. Because the weighing pad is a single unit, and would likely be wired to the computer through a single cord, transfer of the product from bed to bed would be quite simple. The ability to place the pad directly under the mattress substantially decreases the overall load. Since accuracy is dependent on the size of the load, this could aid in achieving the accuracy requested by the client. Also these weighing pads are designed for large loads, often up to 400 pounds, meaning the client could use the device with virtually any patient.


The weighing pad design also has a number of disadvantages. In our research to date we have been unable to find a weighing pad that is capable of obtaining the accuracy requested by the client. Many of these pads on the market are only accurate to .1 of a pound, whereas the client has requested accuracy to .02 to .05 pounds. Also many of the weighing pads are not large



enough to cover the entire area of the hospital bed. With the pad being placed under the mattress, there is the chance that the patient's movement would throw off the instantaneous readings. The placement of the pad under the mattress could lead to patient discomfort which, by the clients request, is unacceptable. Weighing pads are also quite expensive to replace if anything were to go wrong with the design.

### *Tension Meter*

Another alternative design that would be able to record the weight of a patient during the course of the night is the use of a digital tension meter. A digital tension meter measures the precise tension in a cable. The cable is run through prongs which sense the deformation in the cable. The tension meter then transforms this deformation into a digital reading which is then displayed. Our third design incorporates two digital tension meters, steel cable, and a platform upon which the bed rests.



In our design, we would use four digital tension meters (see figure 3) to record the change in weight of the patient. The use of multiple tension meters would distribute the weight

of the person and bed between four separate meters, giving us a more accurate reading. The tension meters that we would like to use would accurately measure the tension in the cable to .01 lbs of force. Digital calibration will be necessary to ensure that only the weight of the patient will be taken. Instead of having the digital tension meter display the readings on the meter screen, many digital tension meters are readily able to interface to computers with the use of a USB port. The digital serial data produced by the digital tension meter will then be directed into the computer with the use of WinWedge Pro. WinWedge captures serial data, custom tailors it, then transfers the data to a PC application such as Excel. Using this specific type of meter, we would be able to manipulate the data-taking interval with WinWedge Pro and observe the data with the use of Excel. This type of digital meter will be allowed to record continuously while connected to a power supply. In our design, we will make use of four of these digital tension meters which will both be separately interfaced to a computer. With the use of Excel, we will be able to take the four different readings and combine them to produce the desired reading.

These tension meters will be attached to steel cables. The steel cable we will use in our design is 3.18 millimeters thick with a max breaking force of 920 pounds, sufficient for the load placed on these cables. These cables will first attach to a platform underneath the bed (refer to figure 3) and then to respective rods which will be anchored to a second platform. The bed will be firmly attached to the platform to avoid any sort of movement.

There are reasonable advantages to this design. All the parts to the design are readily made and obtainable to assemble. We would not need to create an amplifier or A/D converter to manipulate the signal. The tension meter already converts this signal for us. Also, the digital tension meter has a setting for automatic calibration when the size of the wire is known. This would save us time trying to calibrate our design for the patient. Finally, this design has no

irregular discomfort associated with it. Since the device is used under the bed, the patient will experience the same comfort he or she would experience without this device.

There are, however, some disadvantages to this design. One of the difficult obstacles to overcome is the price of this design. Accurate tension meters needed for this design are very expensive. The cost of a single digital tension meter can run anywhere from \$750 to \$1100. We would need to purchase four of these digital tension meters for this design. Another problem associated with this design is the accuracy. If a person moves around significantly at night, the readings from the digital tension meter will show this tension fluctuation. This noise could make analyzing the data very difficult. Finally, the portability and availability of space is a concern. Our client recommended that the device be portable in a window of one to two hours time. This may be possible with design. However, the design is bulky compared to the other two which would make movement difficult with only one person. Also, storage and usage of this design would require more space in the hospital room than the other two designs. This is not a parameter that our client proposed, as there is ample space in the hospital room.

## **Final Design**

Our group analyzed each of the three designs and ranked them according to several categories that we and our client feel are most important to the success of the device. Each design was ranked in its fulfillment of five categories. Rankings between 1 and 5 were given out in each category with 1 being poor and 5 being good. The values given in each category represent a comparison among the three designs. They do not have a quantitative value, as this would require more specific experimental data. Table 1 shows a design matrix and provided each design and the ranking it received in each category. After each design had been ranked in

every category, the total ranking of all five categories was found. Each category was ranked equally because no single aspect of the design is more important than the others.

	Load Cell	Bed Pad	Tension Meter
Cost	3	3	1
Portability	3	5	1
Accuracy	5	1	3
Complexity	3	5	1
Sleep Hindrance	5	1	5
Total	19	15	11

**Table 1 – Design Matrix**

After observing the results of these rankings, our group has decided to move forward with the load cell design. It has the most advantages of any of the designs and our group feels as though it is the design that has the most probability of someday producing a functional device. This design’s high accuracy and versatility are what most influenced our decision to proceed with it. As previously stated, load cells are the most current and efficient way to obtain weight readings. As load cell technology changes we would be able to easily incorporate a new load cell into our design, whereas other scales may just become obsolete. Load cells would be the most probable way for us to achieve the accuracy we need. Accuracy goes down as the size of the load goes up, but our load will be distributed over four separate load cells. This will limit the amount of weight on each individual cell, allowing for a much higher level of accuracy.



## **Future Work**

The next step will be purchasing a load cell and the analog/digital converter so that we can begin to set up the interface between the load cell and the computer. Initially we will purchase a single load cell with a smaller load capacity. This will not produce a functioning prototype but instead, will allow us to set up the interface with minimal cost. In addition, the load cell will most likely require some sort of amplification device to produce readings that are large enough to be recognized by the analog to digital converter. We will use materials available in the BME labs at ECB to begin building an amplifier. We will also need to make sure all four load cells are capable of working in conjunction to produce a single weight readout for the system as a whole.

Power will be provided to the load cell through a battery. A simple 9-volt battery will most likely be sufficient for the small load cell that we will use in our initial prototype. However, a larger more complex power source may be required for a load cell capable of dealing with a greater load.

Once we have set up the initial interface we will need to perform tests to ensure that we can obtain weight readings at chosen intervals. This data must be sent to the computer and imported into computer software that will allow for its manipulation. Microsoft Excel is what we will attempt to use in our prototype. We will also need to test that the readings we are receiving are accurate. This could be done by placing known weights on the scale and observing the read outs.

We will also need to make sure all four load cells are capable of working in conjunction to produce a weight readout for the system as a whole.

If we are able to set up a working interface we must then look in to purchasing load cells that are capable of handling the loads they will see in actual use in the sleep lab. This will require research and perhaps contacting companies to ensure that the load cell will meet all of the requirements.

### **Problems and Possible Resolutions**

The problem we are most likely to face is cost. If we are able to set up an interface that meets our requirements, the load cells and other materials needed to produce a functioning model could end up costing in the range of thousands of dollars. Our best chance to avoid this problem would be to contact companies and explore the possibility of donations or student or university discounts. Without some sort of price reduction, it is very unrealistic that we would be able to obtain the materials at a reasonable cost.

## References

1. "Introduction to Load Cells". Omega Engineering Technical Reference. 10 October 2006 < <http://www.omega.com/prodinfo/loadcells.html> >
2. Harris, Tom. "How Amplifiers Work." How Stuff Works. 12 October 2006 < <http://electronics.howstuffworks.com/amplifier.htm> >
3. Staller, Len. "Understanding analog to digital converter specifications". Embedded Systems Design. 13 October 2006 < <http://www.embedded.com/showArticle.jhtml?articleID=60403334> >
4. Tanaka M, Itoh K, Abe S, Imai K, Masuda T, Koga R, Itoh H, Konomi Y, Kinukawa N, Sakata T. "Irregular patterns in the daily weight chart at night predict body weight regain." Experimental Biology and Medicine. 229:940-945, 2004.
5. Horvath, Tamas. "Obesity and Insomnia Linked by Excitability of Brain Cells." Cell Metabolism. Vol. 1, Issue 4 (April 2005).

## **Product Design Specification**

### **Periodic measurement of nighttime weight change while asleep**

Updated: October 15, 2006

#### **Team Members:**

- Ben Engel: Communications
- Joe Ferris: BWIG
- Kyle Herzog: Leader
- Eric Printz: BSAC

#### **Problem Statement:**

Dr. Juergens would like to be able to measure this weight change accurately on people who are in hospital beds, and be able to view and record the weight changes over various intervals through the night (such as weighing a person who agrees to lay in bed all night at intervals of every 10-15 minutes). These weight readings should be sent to some digital storage software such Microsoft Excel. This is important from a research perspective on numerous potential variables that may affect this weight change, such as medications as well as sleep stage changes. He would also use this to potentially track and eventually predict those at higher risk of developing obesity, correlate this with body fat measurements, activity, and assess for metabolic changes when he does things to improve/correct sleep problems. There is popular concern in the most recent sleep literature about various sleep stages, insomnia, and their impact on obesity. This scale would likely have to support a hospital bed, which sits on 4 legs, and read at an accuracy ideally around 0.02- 0.05 lbs (many scales on the market record to such accuracy, the more accurate the better). Ideally it would be a device which could be moved to a different bed without too much difficulty (such as in 1-2 hours time), and without too complex recalibration.

#### **Client Requirements:**

- Accurate to 0.05 lbs
- No discomfort to patient
- Quick (1-2 hours) transfer time between beds
- 300 lb maximum weight
- Price range of \$300 - \$500
- Provide digital readout of data to computer software, preferably Microsoft Excel

#### **Design Requirements:**

##### **1. Physical and Operational Characteristics**

*a. Performance Requirement:* The device must be operational for up to 12 consecutive hours. The maximum usage would be 12 hours in a 24 hours period.

*b. Safety:* The device should not provide any discomfort to the patient. It should be stable, meaning it should not be possible for it to malfunction and cause the patient any injury.

*c. Accuracy and Reliability:* The device must be accurate to 0.05 lbs. It should be able to obtain readings for a 200 to 300 lb patient, also taking into account the weight of the bed. Readings should be taken at least every 10-15 minutes.

*d. Life in Service:* The device must be operational for 12 hours. It must also be capable of repeated use on consecutive day. The life expectancy will depend upon the durability of the parts being used in the device but it should not need replacement within the next 10 years.

*e. Operating Environment:* The device will be used in a bedroom type setting in the VA hospital (Department of Veterans Affairs). There should be no harmful conditions it undergoes.

*f. Ergonomics:* The device must be comfortable for the patient. The programming aspect should be simple enough training of no more than several minutes is required. It should also be capable of being moved within 1 to 2 hours.

*g. Size and Shape:* The device must be of minimal size. No part of it should protrude from either the top or bottom of the bed. Its interaction with the bed must minimize any sleeping distractions.

*h. Weight:* The device must be a minimal weight so that transfer is not difficult (maximum 20 kg).

*i. Materials:* Materials should minimize cost, discomfort, noise, and visibility.

*j. Aesthetics, Appearance, and Finish:* The device should be appropriate for a hospital setting. It should not interfere with the patients ability to sleep comfortably.

## **2. Product Characteristics:**

*a. Quantity:* One device is required.

*b. Target Product Cost:* The prototype should cost less than \$500 to build.

## **3. Miscellaneous:**

*a. Standards and Specifications:* The device should comply with all regulations established by the FDA for medical instruments. More information can be found on the FDA website.

*b. Customer:* The typical customer would be doctors researching sleep. We are designing it more for this specific study, as opposed to producing it for mass production.

*c. Patient-related concerns:* The patient must feel comfortable, and the device must not prevent sleep in any way.

*d. Competition:* No similar device currently exists.