

Product Design Specifications

December 12, 2007

Title: Accessible Incontinence Control Device

Client: Dr. John Enderle

Team Members: Arin Ellingson, Co-Team Leader
David Schurter, Co-Team Leader
Marty Grasse, Communications
Ben Schoepke, Website Coordinator
Jon Sass, Biomedical Student Advisor Committee Representative

Advisor: Mitch Tyler
UW-Madison Biomedical Engineering

Problem Statement (Adapted from RERC on AMI project description):

Patients with incontinence are unable to control urine flow due to specific disease pathology, trauma, or other causes. Incontinence affects men and women, occurs more frequently with age, and can cause infection, skin irritation, and embarrassment. It negatively affects quality of life and many incontinent patients avoid activities in public, for instance due to the potential for a spastic bladder to spontaneously cause release of urine without warning.

The aim of the project is to design a device to allow the patient (or caregiver) to control and manage urine flow. It should:

- 1) Be easily used by a patient with disabilities
- 2) Allow emptying of the bladder when desired
- 3) Prevent urine flow when not desired
- 4) Provide an indication of the status of the bladder
- 5) Remain indwelling for up to 30 days with no adverse tissue reaction or material degradation

Design Requirements

1. Physical and Operational Characteristics

a. *Performance requirements:* The device must enable control and management of urine flow by patients with disabilities, be able to be switch flow on and off, and provide an indication of the status of the bladder. It must also be able to remain indwelling for up to 30 days without adverse patient reaction or material degradation. The design is

based on a traditional Foley catheter design, with several modifications for improved accessibility and performance.

b. *Safety*: Because the device will be inserted into the urethra and remain indwelling for some time, it must be made of materials that will not cause a reaction or infection in the patient. Foley catheters are typically made of Latex or silicone. According to a urologist, if the device malfunctions and the bladder reaches over-pressures, urine will leak around the catheter before causing damage.

c. *Accuracy and Reliability*: The device will be used to monitor the status of the bladder. This information will be used to assist the patient and caregivers in flow control.

d. *Life in Service*: The device must be able to remain indwelling for up to 30 days before replacement, at which point the Foley catheter and pressure transducer will be discarded. The control circuit, battery pack, and pinch valve should remain viable indefinitely.

e. *Shelf Life*: The device will be sterilized prior to use and should have a shelf life of 5 years, as this is standard for urinary catheters.

f. *Operating Environment*: During use, the device will be inserted through the urethra and into the bladder. It will be in prolonged contact with urine, with typical pH values varying between 4.5 and 8.0.

g. *Ergonomics*: The device must be accessible for patients with disabilities. The flow control mechanism must be easy to use and the bladder status indication must be simple and clear.

h. *Size and Weight*: Traditional Foley catheters range in size from 8 French to 24 French. The device should not be obtrusive or interfere with the patient's daily activities. The control circuit is 5 inches by 2.5 inches by 1.25 inches and is belt-mounted.

i. *Materials*: Materials used must not cause any reaction or infection in the patient. Also, the device must be able to remain indwelling for 30 days with no material degradation. Foley catheters are traditionally made of Latex or silicone

j. *Aesthetics, Appearance, and Finish*: The device should have a smooth and lubricious finish for easy insertion. The control mechanism should be small and the user interface should be simple and easy to operate.

2. Production Characteristics

a. *Quantity*: At this stage, a single prototype device will be designed and produced.

b. *Target Product Cost*: The total cost to design the device must not exceed \$2,000.

3. Miscellaneous

a. *Standards and Specifications:* The device must comply with established industry standards for Foley catheters and other urological medical devices.

b. *Customer:* The device should be simple to use and cater to patients with disabilities such as hearing or vision loss, spinal injuries, or limited movement.

c. *Competition:* Based on our initial research, there are currently no devices on the market that include bladder status monitoring coupled with a patient-controlled valve.