

Digital traction with Japanese finger sleeves

Progress Report 5

Client: Mr. Pape Samb

Advisor: Dr. Justin Williams

Date: 10/09/2025

Team:

Ilia Mikhailenko	imikhailenko@wisc.edu (Co-leader)
Nathan Hansen	ndhansen@wisc.edu (Co-leader)
Nathan Klauck	nklauck@wisc.edu (BSAC)
Ben Willihnganz	bwillihnganz@wisc.edu (BWIG)
Mariamawit Tefera	mstefera@wisc.edu (Team Communicator)
Sam Dudek	dudek4@wisc.edu (BPAG)

Problem Statement

Design a device that allows for precise digital control and support of the components of the human hand through the use of universal buckle strap design and will be mounted onto a stand that can be utilized both on a bed clamp design and a wheeled stand design. This device will be designed to allow for controlled and stable traction during relevant procedures so that proper positioning of the hand can be attained with minimal manual effort.

Brief Status Update

The team has received an update from the client regarding what they are looking for in the design and the specific features the team should prioritize this semester. The team is working to finalize design matrices and prepare for the preliminary presentation this Friday.

Weekly/Ongoing Difficulties

- None to report currently

Current Design

- None to report currently

Materials and Expenses

- None to report currently

Summary of Past Week Accomplishments

- Ilia Mikhailenko
 - Wrote and reviewed the preliminary report, mainly focusing on the conclusion and design matrix portions with contributions to the introduction portion as well
 - Communicated with the team to ensure everything was submitted in a timely manner
- Nathan Hansen
 - Worked on the preliminary design and fabrication section of the preliminary report
 - Developed a hybridized design for the mechanical component on the project
- Nathan Klauck
 - Met with BSAC exec to discuss new training form
 - Conducted research relevant to the project, as well worked on preliminary report
- Ben Willihnganz
 - Worked on the design and material portion of our preliminary report
 - Performed research on materials required for mechanical design
- Mariamawit Tefera
 - Worked on the existing designs, summarizing pds, discussion of testing sections of the preliminary report.
 - Researched more on testing procedures we can follow and use.
- Sam Dudek

- Worked on the motivation, problem statement, biology and physiology, relevant research, and client information for our preliminary report

Upcoming Team and Individual Goals:

The current team goals include furthering individual research and considering design ideas to move forward with.

- Ilia Mikhailenko
 - Develop a plan for bringing this device to life and getting some physical samples
 - Work with the team to potentially split our focus into more of a two team approach, with one focusing on the sleeves and the other focusing on the
- Nathan Hansen
 - Conduct research on potential materials for our chosen design
 - Work on ordering materials to begin prototype fabrication
- Nathan Klauck
 - Prepared for BSAC meeting
 - Start working on designs
 - Further research
- Ben Willihnganz
 - Begin looking for materials in Makers Space and Design Lab
 - Look into potential areas of concern in our initial design ideas
 - Begin first steps of fabrication
- Mariamawit Tefera
 - Finalized preparation for testing and refine our material choice
 - Start ordering materials and confirm where we will find the funding for that.
- Sam Dudek
 - Begin first steps of fabrication
 - Research cost effective materials for our design
 - Look into cost of materials and begin purchasing parts for our prototype
 - Communicate with client about budgeting and if they can buy it for us.

Activities Timesheet




Team Member	Time for the Week	Total Time for the Semester
Ilia Mikhailenko	4	21
Nathan Hansen	5	21
Nathan Klauck	5	18
Ben Willihnganz	3	17
Mariamawit Tefera	3	19
Sam Dudek	4	18

Preliminary Project Timeline:

[illegible]

Design Matrices:

Finger Sleeves

Design Criteria	Weaved Design		Adjustable Velcro Design		Hand Immobilizer	
						
Safety (25)	5/5	25	4/5	20	2/5	10
Ease of Fabrication (20)	2/5	8	5/5	20	4/5	16
Cost (20)	3/5	12	5/5	20	2/5	8
Ease of Use (15)	5/5	15	4/5	12	3/5	9
Comfort (10)	4/5	8	3/5	6	5/5	10
Reusability (10)	2/5	4	4/5	8	2/5	4
Total (100)	72/100		86/100		57/100	

Criteria 1: Cost (25)

- The cost of the design reflects both the raw material price and the overall affordability of manufacturing. Since hospitals and clinics often need multiple devices, keeping costs low is a priority.
- **Design 1: Weaved Design** — This design scores well on cost. The woven material is inexpensive and widely available, but it requires shaping and tailoring into the correct size and form, which adds moderate expense.
- **Design 2: Adjustable Velcro Design** — This design scores the highest because Velcro and supporting materials are cheap, standardized, and easy to produce in bulk. Minimal shaping or customization reduces overall costs.

- **Design 3: Hand Immobilizer** — This design is more expensive due to its layered foam and flexible metal structure. The additional material and complexity make it less cost-effective than the other two options.

Criteria 2: Reusability (25)

- **Reusability** refers to the number of times a device can be applied, cleaned, and maintained before it becomes unsafe or unusable. This is important to reduce waste and maximize long-term value.
- **Design 1: Weaved Design:** This design scores lowest in reusability. Its woven material can fray, stretch, or lose elasticity over repeated uses and sterilizations, reducing durability over time.
- **Design 2: Adjustable Velcro Design:** This design has the highest reusability. Its straps can be adjusted to fit multiple patients, and the overall construction is durable enough to withstand repeated cleaning and use. The main drawback is eventual Velcro wear, though replacement is simple.
- **Design 3: Hand Immobilizer:** This design has moderate reusability. While it is sturdy, the foam padding degrades with repeated cleaning and use, limiting its long-term durability.

Criteria 3: Safety (traction) (15)

- The safety of the design depends on the overall safety for both the patient and physician. This includes both the level and method of traction and dispersion of forces. The weight of this criterion is because safety of all parties needs to be included, as there are procedures being done.
- **Design 1:** This model, which is already commonly used, disperses the force throughout the finger due to its weaved material and keeps the finger steady by use of gravity.
- **Design 2:** This model has two velcro straps which could center too much force on certain parts of the finger, and velcro could eventually wear out.
- **Design 3:** This model has more pressure points around the hand and wrist, which could lead to patient discomfort considering most of the procedures using digital traction are small hand/wrist fractures.

Criteria 4: Ease of Fabrication (15)

- **Ease of Fabrication** refers to the ease of taking available parts and compiling them into our desired product. This was rated in the middle for importance as although any easily fabricated design is important for prototyping and possible future production, the designs cost and reusability play a larger role in the validity of the product.
- **Design 1:** This design requires a multiple size approach and thus requires individual construction for each. Additionally, the conversion of the mess into a cone like shape for application isn't easily achievable, and thus was rated the lowest of the designs
- **Design 2:** This design involves a simplistic manipulation of the material used with a few minor parts and was thus rated second highest for its category

- **Design 3:** This design requires only a foam layer and a flexible metal material beneath, or can be purchased in a similar form and was thus rated the highest for ease of fabrication

Criteria 5: Ease of Use (10)




- The ease of use both depends on the ease of applying the finger sleeve to the patient, but also its ease of use in a multitude of different medical procedures. The weight of this criterion is relatively small because this is already the more simple part of the design, and once it is on, there is not much adjustment needed.
- **Design 1:** This design easily slips onto the finger and grasps back on utilizing gravity and its weaved nature, provides tension and leaves space for operation on the finger, hand, and wrist.
- **Design 2:** This design, although it has no restrictions in procedures it can be used in, requires manual tightening and adjusting by the physician.
- **Design 3:** This design is very easy to strap in initially, but would be very difficult to use in certain applications. This includes castings, any wrist procedure, and more.

Criteria 6: Imaging Compatibility (10)

- Imaging Compatibility refers to how easy the design is to construct with components that are compatible with basic imaging techniques like MRI. This was rated the lowest as although this feature is valuable, it isn't necessary for our prototype.
- **Design 1:** This design involves a completely MRI safe design, with the metal hook being easily swapped for a compatible metal if necessary and was thus rated the highest.
- **Design 2:** This design is also MRI compatible with the only portion at risk being the supporting metal hook. The only reason this was rated slightly lower is due to the difficulty to switch the hook for a compatible metal if necessary, making it slightly less compatible
- **Design 3:** This design requires a flexible metal attached to the back of a foam and thus is not ease to replace and requires a more specific compatible metal and was thus rated the lowest
-

Mechanical/ Frame

Design Criteria	Standing Platform	Bed Clamp & Restraint	Extension Brace
-----------------	-------------------	-----------------------	-----------------

						
Ease of use (25)	4/5	20	5/5	25	3/5	15
Cost (20)	4/5	16	5/5	20	2/5	8
Reusability (20)	4/5	16	5/5	20	2/5	8
Safety (15)	3/5	9	4/5	12	3/5	9
Ease of Fabrication (10)	5/5	8	4/5	4	2/5	10
Versatility (10)	4/5	8	3/5	6	2/5	4
Total (100)	81/100		87/100		54/100	

Criteria 1: Ease of use (25)

- One of the sole purposes of our design is to reduce manual labor and effort a surgeon puts forth in surgery. For this reason, the ease of operation quickly and effectively is the most important aspect of our mechanical design.
- Design 1: The Standing Platform design is effective in ease of use because of its simplicity. This design has very few moving parts and the majority of the mechanism is put inside of the structure itself, but the wheels and multi-use aspect of the design could cause confusion and take away valuable time preparing for procedures and making adjustments.
- Design 2: The Bed Clamp and Restraint design is the most effective in ease of use because of its similarity to preexisting devices. It is/ common for surgical equipment to clip or screw onto the side of the operating table, so most surgeons and residents will be familiar with this concept. This design is also very easily adjusted making fine tuning simple and efficient.

- **Design 3: The Extension Brace** receives the lowest score in ease of use because of the manual effort it takes to apply the compression forces and difficulty in getting the patient in the device. This design also almost entirely covers up either the front or the back of the hand making procedures on the sides of the wrist more difficult and clunky.

Criteria 2: Cost (20)

- The cost of the design reflects raw material price, manufacturing complexity, and overall affordability for hospitals to purchase and use.
- **Design 1: Standing Platform** — Moderate cost. Materials are not expensive, but added wheels and control components raise manufacturing costs slightly.
- **Design 2: Bed Clamp & Restraint** — Lowest cost. Simple design with standard components makes it inexpensive to produce and assemble.
- **Design 3: Extension Brace** — Highest cost. Requires more complex parts, precision hardware, and replaceable padding, increasing expense.

Criteria 3: Reusability (20)

- Reusability is important because hospitals and surgical centers require devices that can withstand repeated sterilization and use across multiple patients. Also, the more durable and easier to clean the device, the more cost-effective and sustainable it is in practice.
- **Design 1: The Standing Platform** design has moderate reusability. While the solid structure and plastic components can endure repeated use, the wheels and moving parts may wear down with time and constant sterilization, which reduces its long-term durability compared to simpler designs.
- **Design 2: The Bed Clamp and Restraint** design scores the highest in reusability. Its simple construction, lack of fragile moving parts, and ability to be easily cleaned and sterilized make it highly suitable for long-term repeated use in clinical settings.
- **Design 3: The Extension Brace** has the lowest reusability. The manual compression components and high-contact areas with the patient's hand and wrist may degrade more quickly with repeated sterilization, while the complex setup makes it more prone to wear out.

Criteria 4: Safety (15)

- The safety of this design depends on the safety for both the patient and physician. This includes its stability, dispersion of forces, and more. The weight of this criterion has to do with safety being relevant for all parties involved, every physician and patient who comes into contact with the device.
- **Design 1: This design** disperses the forces well by using gravity on the hand. Its problem is stability, it is possible that the wheels could move during procedure, and if there is too much tension, the control box could fail and disconnect from the mounting post.
- **Design 2: This design's** clamp at the bottom is its only clamp for stability. This could become a problem if the patient's arm becomes too heavy or the tension provides too much force, leading to failure.
- **Design 3: This design's** main problem happens to be its dispersion of forces. It has too much contact with the patient's hand and wrist, which is precisely the site of most of the procedures that need digital traction.

Criteria 5: Versatility (10)

- The versatility of the design depends on its ability to be used in a multitude of different operations or procedures. The weight of this criterion has to do with the fact that our client has stated the need to use it for multiple procedures.
- Design 1: This design, with its addition of wheels and multiple points of adjustment, is very versatile. Its one drawback is its lack of adjustment to the angle of attachment at the very top.
- Design 2: This design has many areas to adjust the patient's arm which allows it to have many different applications, but is limited by its need to lock directly to a hospital bed.
- Design 3: This design gets a lower score because although it applies tension, it does not leave ample room for different operations. This includes things like casting, where they need space to get around the entire wrist.

Criteria 6: Imaging Compatibility (10)

- The Imaging Compatibility of the designs is graded on how adaptable they are to being used in different imaging devices, specifically MRI compatibility.
- Design 1: The Standing Platform design is fairly suitable for imaging. The body of the device could primarily be built out of a strong plastic material that is good for imaging. In addition to this, the ability for the device to be used standing, sitting, or laying down, allows for numerous possibilities as far as different imaging devices.
- Design 2: The Bed Clamp and Restraint design struggles as far as imaging goes. Its metal frame and body could be made of MRI safe metals, but the lack of adjustability and mobility reduces many imaging possibilities.
- Design 3: The Extension Brace is most suitable for imaging. This device is unique in that it can stand alone as it does not rely on gravity for the application of tension. This allows the device to be directly inserted into different hand, wrist, and arm imaging devices. The lack of a “clunky” body helps this device tremendously.